



NEOTEMP 1910

Neonatal Skin Temperature Sensor



For use with
Dräger Medical/Air Shields
Isolette Incubators
IICS Models C-100, C-450, C-550 &
PM-78 Infant Radiant Warmers

[Replaces Air Shields Preemie Probe 1
#68-209-20 and #68-209-30 &
Dräger Medical #MU06933 & #MU06935]

- Temp Accuracy** NovaMed baby-sensitive thermistor sensor guarantees temperature reading sensitivity with a stable, highly accurate measurement. Molded plug assembly provides direct connection to the neonate/incubator/infant warmer, eliminating the need for interface cable.
- Sensor Safety** Flexible 60" lead wire facilitates sensor placement on the infant eliminating the need for an interface/interconnect cable and further ensuring precision and safety.
- Ease of Use** Packaged individually in a Tyvek pouch with an insulated adhesive Mylar temperature probe cover (one inch circular disk).
- Gentle** Neonatal sensitive hypoallergenic adhesive/reflective Mylar probe cover provides insulation from heat and external sources while ensuring precise temperature monitoring capabilities.

NEOTEMP Neonatal Skin Temperature Sensor Disposable with Insulated Mylar Disc



- Designed for Dräger Medical/Air Shields C-100, C-450, C-550 Isolette Infant Incubators & PM-78 Infant Radiant Warmers.
- Replaces Air Shields Preemie Probe1 #68-209-20 & #68-209-30 & Dräger Medical #MU06933 & # MU06935 (Reusable #68-209-70, Dräger Medical #MU06951).
- NEOTEMP 1910; Packaged 20/box, Order Ref. #10-1910-020
NEOTEMP 1910; Packaged 60/case, Order Ref. #10-1910-060.
- 3 Pin Din Connector with molded thermistor Flex-reinforcement.